

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121224\
 Data File : VX044250.D
 Acq On : 12 Dec 2024 14:23
 Operator : JC/MD
 Sample : P5261-08 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-01-20241209

Quant Time: Dec 13 01:40:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	109125	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	206340	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	178333	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	69576	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	95018	49.661	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.320%
35) Dibromofluoromethane	5.379	113	68518	47.945	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.900%
50) Toluene-d8	8.647	98	250910	52.041	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.080%
62) 4-Bromofluorobenzene	11.079	95	78938	46.856	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	93.720%
Target Compounds						
					Qvalue	
44) Trichloroethene	7.129	130	439	0.314	ug/l	56

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121224\
Data File : VX044250.D
Acq On : 12 Dec 2024 14:23
Operator : JC/MD
Sample : P5261-08 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-01-20241209

Quant Time: Dec 13 01:40:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

