

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121224\
 Data File : VX044260.D
 Acq On : 12 Dec 2024 18:15
 Operator : JC/MD
 Sample : P5261-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GM3-RW3-MW1-20241210

Quant Time: Dec 13 01:49:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	116439	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	225040	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	194438	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	79574	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	104886	51.375	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 102.760%	
35) Dibromofluoromethane	5.379	113	74979	48.106	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 96.220%	
50) Toluene-d8	8.647	98	266458	50.674	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.340%	
62) 4-Bromofluorobenzene	11.079	95	90711	49.370	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 98.740%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.307	101	1951	1.373	ug/l	95
16) Acetone	2.386	43	1376	1.653	ug/l #	74
20) Methylene Chloride	2.782	84	564	0.361	ug/l #	60
30) Chloroform	5.093	83	1304	0.425	ug/l #	78
44) Trichloroethene	7.123	130	26062	17.085	ug/l	92
64) Tetrachloroethene	9.269	164	2174	1.685	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121224\
Data File : VX044260.D
Acq On : 12 Dec 2024 18:15
Operator : JC/MD
Sample : P5261-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GM3-RW3-MW1-20241210

Quant Time: Dec 13 01:49:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

