

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121225\
 Data File : VX048846.D
 Acq On : 12 Dec 2025 12:04
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Dec 12 22:11:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 05 03:27:34 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	161108	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	337027	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	361638	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	186868	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	122499	56.650	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.300%
35) Dibromofluoromethane	5.373	113	109641	47.248	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.500%
50) Toluene-d8	8.629	98	386814	47.553	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.100%
62) 4-Bromofluorobenzene	11.061	95	165924	59.154	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	118.300%
Target Compounds						
						Qvalue
27) cis-1,2-Dichloroethene	4.477	96	1333	0.619	ug/l	98
44) Trichloroethene	7.117	130	896	0.364	ug/l	67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121225\
Data File : VX048846.D
Acq On : 12 Dec 2025 12:04
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Dec 12 22:11:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 05 03:27:34 2025
Response via : Initial Calibration

