

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121225\
 Data File : VX048847.D
 Acq On : 12 Dec 2025 12:26
 Operator : JC/MD
 Sample : Q3828-05DL 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RMW-03B-90-120925DL

Quant Time: Dec 12 22:11:32 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 05 03:27:34 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.531	168	149273	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	310587	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	339410	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	178609	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	116556	58.176	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 116.360%	
35) Dibromofluoromethane	5.373	113	101503	47.464	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 94.920%	
50) Toluene-d8	8.635	98	344261	45.925	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 91.860%#	
62) 4-Bromofluorobenzene	11.061	95	150309	58.149	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 116.300%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.380	62	876	0.429	ug/l #	88
9) 1,1,2-Trichlorotrifluo...	2.337	101	176276	103.794	ug/l	98
12) 1,1-Dichloroethene	2.331	96	1043	0.602	ug/l	84
24) 1,1-Dichloroethane	3.605	63	8309	2.740	ug/l	96
27) cis-1,2-Dichloroethene	4.477	96	170392	85.352	ug/l	94
44) Trichloroethene	7.110	130	139882	61.742	ug/l	96
64) Tetrachloroethene	9.263	164	705	0.285	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121225\
Data File : VX048847.D
Acq On : 12 Dec 2025 12:26
Operator : JC/MD
Sample : Q3828-05DL 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RMW-03B-90-120925DL

Quant Time: Dec 12 22:11:32 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 05 03:27:34 2025
Response via : Initial Calibration

