

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121225\
 Data File : VX048848.D
 Acq On : 12 Dec 2025 12:48
 Operator : JC/MD
 Sample : Q3835-01DL 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-11A-13.5-121025DL

Quant Time: Dec 12 22:11:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 05 03:27:34 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	143724	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	302254	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	330363	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	169799	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	116888	60.594	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 121.180%#		
35) Dibromofluoromethane	5.373	113	99546	47.833	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 95.660%		
50) Toluene-d8	8.628	98	341875	46.864	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 93.720%		
62) 4-Bromofluorobenzene	11.061	95	146324	58.168	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 116.340%		
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.380	62	545	0.277	ug/l #	49
9) 1,1,2-Trichlorotrifluo...	2.331	101	2920	1.786	ug/l	97
12) 1,1-Dichloroethene	2.325	96	441	0.264	ug/l #	62
27) cis-1,2-Dichloroethene	4.477	96	40079	20.851	ug/l	97
44) Trichloroethene	7.110	130	126532	57.389	ug/l	99
64) Tetrachloroethene	9.256	164	533	0.221	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121225\
Data File : VX048848.D
Acq On : 12 Dec 2025 12:48
Operator : JC/MD
Sample : Q3835-01DL 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-11A-13.5-121025DL

Quant Time: Dec 12 22:11:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 05 03:27:34 2025
Response via : Initial Calibration

