

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121225\
 Data File : VX048850.D
 Acq On : 12 Dec 2025 13:31
 Operator : JC/MD
 Sample : Q3828-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BR-02-133-120925

Quant Time: Dec 12 22:12:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 05 03:27:34 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.537	168	139671	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	296647	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	321216	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	171700	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	114631	61.148	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	122.300%#	
35) Dibromofluoromethane	5.373	113	99461	48.695	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	97.380%	
50) Toluene-d8	8.634	98	342777	47.876	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	95.760%	
62) 4-Bromofluorobenzene	11.061	95	147914	59.911	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	119.820%	
Target Compounds						
						Qvalue
44) Trichloroethene	7.104	130	664	0.307	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121225\
Data File : VX048850.D
Acq On : 12 Dec 2025 13:31
Operator : JC/MD
Sample : Q3828-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BR-02-133-120925

Quant Time: Dec 12 22:12:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 05 03:27:34 2025
Response via : Initial Calibration

