

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121225\
 Data File : VX048853.D
 Acq On : 12 Dec 2025 14:36
 Operator : JC/MD
 Sample : Q3835-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OWBR-01-170-121025

Quant Time: Dec 12 22:13:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 05 03:27:34 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	137404	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	292602	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	323774	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	171165	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	116026	62.914	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 125.820%#		
35) Dibromofluoromethane	5.373	113	100467	49.867	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 99.740%		
50) Toluene-d8	8.635	98	341625	48.375	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 96.740%		
62) 4-Bromofluorobenzene	11.061	95	148928	61.156	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 122.320%		
Target Compounds						
					Qvalue	
44) Trichloroethene	7.117	130	3482	1.631	ug/l	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121225\
Data File : VX048853.D
Acq On : 12 Dec 2025 14:36
Operator : JC/MD
Sample : Q3835-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OWBR-01-170-121025

Quant Time: Dec 12 22:13:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 05 03:27:34 2025
Response via : Initial Calibration

