

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121225\
 Data File : VX048856.D
 Acq On : 12 Dec 2025 15:42
 Operator : JC/MD
 Sample : Q3849-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OW-07B-65-121125

Quant Time: Dec 12 22:14:32 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 05 03:27:34 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	137566	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	283716	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	314168	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	163467	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	111675	60.483	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 120.960%#		
35) Dibromofluoromethane	5.373	113	97235	49.775	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 99.540%		
50) Toluene-d8	8.635	98	330139	48.212	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 96.420%		
62) 4-Bromofluorobenzene	11.061	95	141764	60.037	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 120.080%		
Target Compounds						
						Qvalue
4) Vinyl Chloride	1.380	62	1038	0.551	ug/l	# 84
12) 1,1-Dichloroethene	2.325	96	138025	86.405	ug/l	99
21) trans-1,2-Dichloroethene	3.093	96	6500	3.987	ug/l	96
24) 1,1-Dichloroethane	3.605	63	21611	7.732	ug/l	97
27) cis-1,2-Dichloroethene	4.477	96	217818	118.394	ug/l	96
42) 1,2-Dichloroethane	6.074	62	11408	4.064	ug/l	99
44) Trichloroethene	7.110	130	44794	21.644	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121225\
Data File : VX048856.D
Acq On : 12 Dec 2025 15:42
Operator : JC/MD
Sample : Q3849-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OW-07B-65-121125

Quant Time: Dec 12 22:14:32 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 05 03:27:34 2025
Response via : Initial Calibration

