

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121225\
 Data File : VX048857.D
 Acq On : 12 Dec 2025 16:03
 Operator : JC/MD
 Sample : Q3849-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RMW-02B-67-121125

Quant Time: Dec 12 22:14:54 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 05 03:27:34 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.537	168	108718	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	223506	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	232553	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	124059	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	83588	57.284	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	114.560%
35) Dibromofluoromethane	5.373	113	74638	48.500	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.000%
50) Toluene-d8	8.634	98	241877	44.838	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	89.680%#
62) 4-Bromofluorobenzene	11.061	95	105624	56.782	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	113.560%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.380	62	363679	244.295	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.337	101	19135	15.470	ug/l	98
12) 1,1-Dichloroethene	2.325	96	136286	107.954	ug/l	95
17) Carbon Disulfide	2.526	76	6690	1.825	ug/l	95
21) trans-1,2-Dichloroethene	3.087	96	12157	9.435	ug/l	88
24) 1,1-Dichloroethane	3.599	63	48241	21.839	ug/l	99
27) cis-1,2-Dichloroethene	4.477	96	2458672	1691.012	ug/l	95
40) Benzene	6.019	78	4436	0.710	ug/l	97
42) 1,2-Dichloroethane	6.074	62	21399	9.676	ug/l	99
44) Trichloroethene	7.116	130	4413361	2706.963	ug/l	92
64) Tetrachloroethene	9.256	164	68900	40.647	ug/l	99
88) 1,4-Dichlorobenzene	12.024	146	2189	0.515	ug/l #	37

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121225\
Data File : VX048857.D
Acq On : 12 Dec 2025 16:03
Operator : JC/MD
Sample : Q3849-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RMW-02B-67-121125

Quant Time: Dec 12 22:14:54 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 05 03:27:34 2025
Response via : Initial Calibration

