

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121225\
 Data File : VX048861.D
 Acq On : 12 Dec 2025 17:31
 Operator : JC/MD
 Sample : Q3848-01DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-18B-57.5-121125DL

Quant Time: Dec 12 22:16:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 05 03:27:34 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	129788	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	267248	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	292684	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	156524	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	97577	56.015	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 112.020%	
35) Dibromofluoromethane	5.373	113	86195	46.842	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 93.680%	
50) Toluene-d8	8.635	98	294325	45.631	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 91.260%#	
62) 4-Bromofluorobenzene	11.061	95	127652	57.392	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 114.780%	
Target Compounds						
4) Vinyl Chloride	1.380	62	10738	6.042	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.331	101	3432	2.324	ug/l	97
12) 1,1-Dichloroethene	2.325	96	1525	1.012	ug/l	87
16) Acetone	2.380	43	1659	2.488	ug/l	96
17) Carbon Disulfide	2.520	76	1059	0.242	ug/l #	38
21) trans-1,2-Dichloroethene	3.087	96	673	0.438	ug/l	92
24) 1,1-Dichloroethane	3.599	63	2411	0.914	ug/l #	82
27) cis-1,2-Dichloroethene	4.477	96	153034	88.166	ug/l	95
44) Trichloroethene	7.110	130	86973	44.614	ug/l	95
64) Tetrachloroethene	9.250	164	991	0.465	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121225\
Data File : VX048861.D
Acq On : 12 Dec 2025 17:31
Operator : JC/MD
Sample : Q3848-01DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-18B-57.5-121125DL

Quant Time: Dec 12 22:16:40 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 05 03:27:34 2025
Response via : Initial Calibration

