

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121225\
 Data File : VX048863.D
 Acq On : 12 Dec 2025 18:14
 Operator : JC/MD
 Sample : Q3826-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-2

Quant Time: Dec 12 22:17:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 05 03:27:34 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.537	168	112139	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	235902	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	269420	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	142484	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	92811	61.664	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 123.320%#	
35) Dibromofluoromethane	5.367	113	80292	49.432	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.860%	
50) Toluene-d8	8.634	98	276756	48.608	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 97.220%	
62) 4-Bromofluorobenzene	11.061	95	120272	61.259	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 122.520%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.343	101	1226	0.961	ug/l	94
16) Acetone	2.373	43	51245	88.938	ug/l	97
18) Methyl Acetate	2.703	43	2035	1.255	ug/l #	86
20) Methylene Chloride	2.788	84	118410	82.218	ug/l	95
25) 2-Butanone	4.556	43	4166	4.899	ug/l #	86
27) cis-1,2-Dichloroethene	4.483	96	1193	0.795	ug/l	60
43) Isopropyl Acetate	6.324	43	7425	1.845	ug/l	97
44) Trichloroethene	7.116	130	8230	4.783	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121225\
Data File : VX048863.D
Acq On : 12 Dec 2025 18:14
Operator : JC/MD
Sample : Q3826-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-2

Quant Time: Dec 12 22:17:26 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 05 03:27:34 2025
Response via : Initial Calibration

