

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121318\
 Data File : VX006492.D
 Acq On : 13 Dec 2018 13:57
 Operator : JC/MD
 Sample : J6358-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-HN-MW271-20181210

Quant Time: Dec 14 09:54:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	646700	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1035387	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	937526	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	438785	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	351127	53.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.38%	
35) Dibromofluoromethane	5.50	113	306544	45.73	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.46%	
50) Toluene-d8	8.71	98	1211883	49.70	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.40%	
62) 4-Bromofluorobenzene	11.13	95	446755	48.62	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.24%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121318\
Data File : VX006492.D
Acq On : 13 Dec 2018 13:57
Operator : JC/MD
Sample : J6358-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-HN-MW27I-20181210

Quant Time: Dec 14 09:54:09 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

