

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121318\
 Data File : VX006493.D
 Acq On : 13 Dec 2018 14:24
 Operator : JC/MD
 Sample : J6350-01DL 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SV-27781LDL

Quant Time: Dec 14 00:43:54 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	643558	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1025752	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	938887	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	444748	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	347442	52.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.78%	
35) Dibromofluoromethane	5.51	113	312203	47.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.02%	
50) Toluene-d8	8.71	98	1202123	49.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.52%	
62) 4-Bromofluorobenzene	11.13	95	448782	49.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.60%	
Target Compounds						
					Qvalue	
16) Acetone	2.45	43	787732	255.499	ug/l	96
25) 2-Butanone	4.70	43	755767	171.798	ug/l	99
29) Tetrahydrofuran	5.15	42	608320	198.966	ug/l	98
68) m/p-Xylenes	10.35	106	9304	0.701	ug/l	97
69) o-Xylene	10.69	106	5403	0.412	ug/l	91
80) 1,3,5-Trimethylbenzene	11.51	105	8466	0.331	ug/l	92
84) 1,2,4-Trimethylbenzene	11.80	105	31578	1.197	ug/l	99
95) Naphthalene	13.83	128	17561	0.514	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121318\
Data File : VX006493.D
Acq On : 13 Dec 2018 14:24
Operator : JC/MD
Sample : J6350-01DL 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SV-27781LDL

Quant Time: Dec 14 00:43:54 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

