

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121318\
 Data File : VX006499.D
 Acq On : 13 Dec 2018 17:04
 Operator : JC/MD
 Sample : PB115614ZHE#07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB115614ZHE#07

Quant Time: Dec 14 01:01:20 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	633395	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1016643	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	946792	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	452503	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	344115	53.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.44%	
35) Dibromofluoromethane	5.51	113	303401	46.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.18%	
50) Toluene-d8	8.71	98	1209882	50.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.06%	
62) 4-Bromofluorobenzene	11.13	95	450350	49.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.82%	
Target Compounds						
16) Acetone	2.45	43	12308	4.056	ug/l	93
18) Methyl Acetate	2.78	43	26826	2.644	ug/l	93
20) Methylene Chloride	2.86	84	10578	1.638	ug/l #	92
43) Isopropyl Acetate	6.46	43	34551	2.161	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121318\
Data File : VX006499.D
Acq On : 13 Dec 2018 17:04
Operator : JC/MD
Sample : PB115614ZHE#07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB115614ZHE#07

Quant Time: Dec 14 01:01:20 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

