

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121318\  
 Data File : VX006502.D  
 Acq On : 13 Dec 2018 18:24  
 Operator : JC/MD  
 Sample : J6199-06  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 18 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 SU-04-121218-A

Quant Time: Dec 14 01:07:41 2018  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X112918W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Nov 30 03:55:33 2018  
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	641453	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1039420	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	959443	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	468111	50.00	ug/l	0.00

## System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.07	65	351232	53.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.28%	
35) Dibromofluoromethane	5.51	113	308943	45.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.82%	
50) Toluene-d8	8.71	98	1227427	50.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.28%	
62) 4-Bromofluorobenzene	11.13	95	458190	49.67	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.34%	

## Target Compounds

					Qvalue	
16) Acetone	2.45	43	45334	14.752	ug/l	94
18) Methyl Acetate	2.78	43	21398	2.082	ug/l	92
20) Methylene Chloride	2.86	84	22974	3.513	ug/l	93
43) Isopropyl Acetate	6.46	43	33543	2.052	ug/l	97
84) 1,2,4-Trimethylbenzene	11.80	105	42901	1.545	ug/l	96
95) Naphthalene	13.83	128	99932	2.779	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121318\  
 Data File : VX006502.D  
 Acq On : 13 Dec 2018 18:24  
 Operator : JC/MD  
 Sample : J6199-06  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 18 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 SU-04-121218-A

Quant Time: Dec 14 01:07:41 2018  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X112918W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Nov 30 03:55:33 2018  
 Response via : Initial Calibration

