

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121318\
 Data File : VX006506.D
 Acq On : 13 Dec 2018 20:10
 Operator : JC/MD
 Sample : J6373-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-A-FRAME

Quant Time: Dec 14 01:15:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	635550	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1032966	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	960858	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	466860	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	350944	54.09	ug/l	0.00
Spiked Amount			Recovery	=	108.18%	
35) Dibromofluoromethane	5.50	113	305064	45.61	ug/l	0.00
Spiked Amount			Recovery	=	91.22%	
50) Toluene-d8	8.71	98	1223390	50.29	ug/l	0.00
Spiked Amount			Recovery	=	100.58%	
62) 4-Bromofluorobenzene	11.13	95	458316	50.00	ug/l	0.00
Spiked Amount			Recovery	=	100.00%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.44	43	16413	5.391	ug/l	96
18) Methyl Acetate	2.78	43	30037	2.950	ug/l #	87
20) Methylene Chloride	2.86	84	1324827	204.450	ug/l	93
43) Isopropyl Acetate	6.46	43	35163	2.165	ug/l	99
95) Naphthalene	13.83	128	342208	9.543	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121318\
Data File : VX006506.D
Acq On : 13 Dec 2018 20:10
Operator : JC/MD
Sample : J6373-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-A-FRAME

Quant Time: Dec 14 01:15:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

