

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121319\
 Data File : VX014014.D
 Acq On : 13 Dec 2019 18:58
 Operator : JC/SP
 Sample : VX1213WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1213WBL01

Quant Time: Dec 14 01:56:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 13 17:10:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	571235	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	868827	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	779833	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	340083	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	314962	48.65	ug/l	0.00
Spiked Amount			Recovery	=	97.30%	
35) Dibromofluoromethane	5.48	113	257138	48.69	ug/l	0.00
Spiked Amount			Recovery	=	97.38%	
50) Toluene-d8	8.71	98	1031070	50.14	ug/l	0.00
Spiked Amount			Recovery	=	100.28%	
62) 4-Bromofluorobenzene	11.14	95	346853	46.15	ug/l	0.00
Spiked Amount			Recovery	=	92.30%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121319\
Data File : VX014014.D
Acq On : 13 Dec 2019 18:58
Operator : JC/SP
Sample : VX1213WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1213WBL01

Quant Time: Dec 14 01:56:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 13 17:10:55 2019
Response via : Initial Calibration

