

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121319\
 Data File : VX014019.D
 Acq On : 13 Dec 2019 21:18
 Operator : JC/SP
 Sample : K6182-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Dec 14 02:50:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 13 17:10:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	587244	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	902309	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	802799	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	335440	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	327364	49.19	ug/l	0.00
Spiked Amount			Recovery	=	98.38%	
35) Dibromofluoromethane	5.48	113	264322	48.19	ug/l	0.00
Spiked Amount			Recovery	=	96.38%	
50) Toluene-d8	8.71	98	1058735	49.58	ug/l	0.00
Spiked Amount			Recovery	=	99.16%	
62) 4-Bromofluorobenzene	11.14	95	353187	45.25	ug/l	0.00
Spiked Amount			Recovery	=	90.50%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121319\
Data File : VX014019.D
Acq On : 13 Dec 2019 21:18
Operator : JC/SP
Sample : K6182-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Dec 14 02:50:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 13 17:10:55 2019
Response via : Initial Calibration

