

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX121319\
 Data File : VX014021.D
 Acq On : 13 Dec 2019 22:04
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 17 03:11:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	99	0.00
2 T	Dichlorodifluoromethane	0.447	0.409	8.5	94	0.00
3 P	Chloromethane	0.600	0.573	4.5	97	0.00
4 C	Vinyl Chloride	0.633	0.591	6.6#	97	0.00
5 T	Bromomethane	0.450	0.387	14.0	92	0.00
6 T	Chloroethane	0.391	0.367	6.1	97	0.00
7 T	Trichlorofluoromethane	0.784	0.727	7.3	95	0.00
8 T	Diethyl Ether	0.368	0.339	7.9	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.473	0.427	9.7	94	0.00
10 T	Methyl Iodide	0.575	0.601	-4.5	97	0.00
11 T	Tert butyl alcohol	0.122	0.123	-0.8	108	0.00
12 CM	1,1-Dichloroethene	0.481	0.447	7.1#	96	0.00
13 T	Acrolein	0.070	0.088	-25.7#	135	0.00
14 T	Allyl chloride	0.859	0.836	2.7	99	0.00
15 T	Acrylonitrile	0.286	0.284	0.7	103	0.00
16 T	Acetone	0.368	0.238	35.3#	64	0.00
17 T	Carbon Disulfide	1.418	1.239	12.6	95	0.00
18 T	Methyl Acetate	0.729	0.737	-1.1	105	0.00
19 T	Methyl tert-butyl Ether	1.580	1.543	2.3	99	0.00
20 T	Methylene Chloride	0.579	0.530	8.5	100	0.00
21 T	trans-1,2-Dichloroethene	0.530	0.486	8.3	97	0.00
22 T	Diisopropyl ether	1.711	1.657	3.2	98	0.00
23 T	Vinyl Acetate	1.397	1.201	14.0	86	0.00
24 P	1,1-Dichloroethane	0.952	0.898	5.7	99	0.00
25 T	2-Butanone	0.454	0.398	12.3	87	0.00
26 T	2,2-Dichloropropane	0.739	0.657	11.1	90	0.00
27 T	cis-1,2-Dichloroethene	0.599	0.564	5.8	98	0.00
28 T	Bromochloromethane	0.332	0.370	-11.4	99	0.00
29 T	Tetrahydrofuran	0.254	0.258	-1.6	104	0.00
30 C	Chloroform	0.903	0.870	3.7#	98	0.00
31 T	Cyclohexane	0.847	0.801	5.4	94	0.00
32 T	1,1,1-Trichloroethane	0.769	0.730	5.1	97	0.00
33 S	1,2-Dichloroethane-d4	0.567	0.549	3.2	98	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	102	0.00
35 S	Dibromofluoromethane	0.304	0.288	5.3	98	0.00
36 T	1,1-Dichloropropene	0.459	0.418	8.9	97	0.00
37 T	Ethyl Acetate	0.507	0.489	3.6	99	0.00
38 T	Carbon Tetrachloride	0.418	0.393	6.0	96	0.00
39 T	Methylcyclohexane	0.566	0.499	11.8	93	0.00
40 TM	Benzene	1.412	1.296	8.2	98	0.00
41 T	Methacrylonitrile	0.281	0.275	2.1	104	0.00
42 TM	1,2-Dichloroethane	0.479	0.443	7.5	98	0.00
43 T	Isopropyl Acetate	0.838	0.813	3.0	101	0.00
44 TM	Trichloroethene	0.390	0.359	7.9	101	0.00
45 C	1,2-Dichloropropane	0.362	0.341	5.8#	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX121319\
 Data File : VX014021.D
 Acq On : 13 Dec 2019 22:04
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 17 03:11:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.239	0.218	8.8	99	0.00
47 T	Bromodichloromethane	0.451	0.427	5.3	98	0.00
48 T	Methyl methacrylate	0.411	0.400	2.7	100	0.00
49 T	1,4-Dioxane	0.008	0.009	-12.5	112	0.00
50 S	Toluene-d8	1.183	1.132	4.3	97	0.00
51 T	4-Methyl-2-Pentanone	0.527	0.517	1.9	102	0.00
52 CM	Toluene	0.892	0.821	8.0#	98	0.00
53 T	t-1,3-Dichloropropene	0.494	0.477	3.4	96	0.00
54 T	cis-1,3-Dichloropropene	0.550	0.533	3.1	97	0.00
55 T	1,1,2-Trichloroethane	0.358	0.333	7.0	98	0.00
56 T	Ethyl methacrylate	0.562	0.553	1.6	99	0.00
57 T	1,3-Dichloropropane	0.602	0.569	5.5	99	0.00
58 T	2-Chloroethyl Vinyl ether	0.213	0.213	0.0	98	0.00
59 T	2-Hexanone	0.424	0.395	6.8	95	0.00
60 T	Dibromochloromethane	0.355	0.352	0.8	99	0.00
61 T	1,2-Dibromoethane	0.366	0.348	4.9	98	0.00
62 S	4-Bromofluorobenzene	0.433	0.409	5.5	96	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	102	0.00
64 T	Tetrachloroethene	0.442	0.505	-14.3	121	0.00
65 PM	Chlorobenzene	1.063	0.974	8.4	97	0.00
66 T	1,1,1,2-Tetrachloroethane	0.380	0.365	3.9	100	0.00
67 C	Ethyl Benzene	1.828	1.702	6.9#	96	0.00
68 T	m/p-Xylenes	0.703	0.651	7.4	96	0.00
69 T	o-Xylene	0.689	0.642	6.8	97	0.00
70 T	Styrene	1.166	1.102	5.5	96	0.00
71 P	Bromoform	0.305	0.299	2.0	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
73 T	Isopropylbenzene	3.520	3.317	5.8	96	0.00
74 T	N-amyl acetate	1.650	1.624	1.6	98	0.00
75 P	1,1,2,2-Tetrachloroethane	1.220	1.151	5.7	96	0.00
76 T	1,2,3-Trichloropropane	1.085	1.091	-0.6	99	0.00
77 T	Bromobenzene	0.969	0.879	9.3	97	0.00
78 T	n-propylbenzene	3.949	3.723	5.7	94	0.00
79 T	2-Chlorotoluene	2.400	2.232	7.0	95	0.00
80 T	1,3,5-Trimethylbenzene	2.963	2.793	5.7	95	0.00
81 T	trans-1,4-Dichloro-2-butene	0.385	0.394	-2.3	98	0.00
82 T	4-Chlorotoluene	2.784	2.550	8.4	95	0.00
83 T	tert-Butylbenzene	2.883	2.483	13.9	88	0.00
84 T	1,2,4-Trimethylbenzene	2.967	2.790	6.0	95	0.00
85 T	sec-Butylbenzene	3.404	3.164	7.1	94	0.00
86 T	p-Isopropyltoluene	3.114	2.905	6.7	94	0.00
87 T	1,3-Dichlorobenzene	1.698	1.501	11.6	94	0.00
88 T	1,4-Dichlorobenzene	1.726	1.503	12.9	94	0.00
89 T	n-Butylbenzene	2.657	2.387	10.2	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX121319\
Data File : VX014021.D
Acq On : 13 Dec 2019 22:04
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 17 03:11:12 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.559	0.524	6.3	93	0.00
91 T	1,2-Dichlorobenzene	1.709	1.535	10.2	94	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.270	0.247	8.5	98	0.00
93 T	1,2,4-Trichlorobenzene	1.111	0.948	14.7	88	0.00
94 T	Hexachlorobutadiene	0.534	0.446	16.5	88	0.00
95 T	Naphthalene	3.263	3.055	6.4	92	0.00
96 T	1,2,3-Trichlorobenzene	1.096	0.996	9.1	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6