

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121322\
 Data File : VX033386.D
 Acq On : 13 Dec 2022 12:49
 Operator : JC/MD
 Sample : N6001-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-3

Quant Time: Dec 14 05:36:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	127886	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	200660	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	177582	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	82332	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	36087	43.985	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.980%
35) Dibromofluoromethane	5.385	113	40283	44.633	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.260%
50) Toluene-d8	8.647	98	83255	44.096	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	88.200%#
62) 4-Bromofluorobenzene	11.079	95	63819	44.207	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.420%
Target Compounds						
16) Acetone	2.386	43	2112	3.479	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121322\
Data File : VX033386.D
Acq On : 13 Dec 2022 12:49
Operator : JC/MD
Sample : N6001-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-3

Quant Time: Dec 14 05:36:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

