

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121322\
 Data File : VX033394.D
 Acq On : 13 Dec 2022 15:54
 Operator : JC/MD
 Sample : N5865-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-01

Quant Time: Dec 14 05:39:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	138117	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	224824	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	201298	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92886	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	44075	49.742	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.480%
35) Dibromofluoromethane	5.385	113	49192	48.646	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.300%
50) Toluene-d8	8.653	98	99716	47.138	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.280%
62) 4-Bromofluorobenzene	11.079	95	77301	47.791	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.580%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	3852	5.875	ug/l	92
20) Methylene Chloride	2.788	84	3239	2.215	ug/l #	88
67) Ethyl Benzene	10.195	91	3300	0.465	ug/l	96
68) m/p-Xylenes	10.305	106	2894	1.047	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121322\
Data File : VX033394.D
Acq On : 13 Dec 2022 15:54
Operator : JC/MD
Sample : N5865-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-01

Quant Time: Dec 14 05:39:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

