

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121322\
 Data File : VX033395.D
 Acq On : 13 Dec 2022 16:17
 Operator : JC/MD
 Sample : N5865-11DL 2000X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-01DL

Quant Time: Dec 14 05:39:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	133182	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	212582	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	189817	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	86382	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	41401	48.456	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.920%
35) Dibromofluoromethane	5.385	113	45710	47.806	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.620%
50) Toluene-d8	8.647	98	94573	47.281	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.560%
62) 4-Bromofluorobenzene	11.079	95	73053	47.766	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.540%
Target Compounds						
27) cis-1,2-Dichloroethene	4.489	96	80900	51.252	ug/l	88
44) Trichloroethene	7.123	130	27097	18.082	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121322\
Data File : VX033395.D
Acq On : 13 Dec 2022 16:17
Operator : JC/MD
Sample : N5865-11DL 2000X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-01DL

Quant Time: Dec 14 05:39:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

