

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121322\
 Data File : VX033398.D
 Acq On : 13 Dec 2022 17:26
 Operator : JC/MD
 Sample : N6001-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-12

Quant Time: Dec 14 05:40:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	122976	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	196491	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	180890	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85330	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	38024	48.197	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.400%
35) Dibromofluoromethane	5.385	113	42448	48.030	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.060%
50) Toluene-d8	8.647	98	89634	48.482	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.960%
62) 4-Bromofluorobenzene	11.079	95	70847	50.117	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.240%
Target Compounds						
					Qvalue	
31) Cyclohexane	5.477	56	3336	1.779	ug/l	85
39) Methylcyclohexane	7.385	83	1362	0.706	ug/l #	78
40) Benzene	6.037	78	64569	12.852	ug/l	99
52) Toluene	8.720	92	2643	0.814	ug/l	100
67) Ethyl Benzene	10.195	91	126688	19.852	ug/l	99
68) m/p-Xylenes	10.305	106	3890	1.567	ug/l	95
69) o-Xylene	10.640	106	1156	0.471	ug/l	81
73) Isopropylbenzene	10.963	105	9912	1.603	ug/l	99
78) n-propylbenzene	11.305	91	13544	1.961	ug/l	99
95) Naphthalene	13.780	128	10233	1.796	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121322\
Data File : VX033398.D
Acq On : 13 Dec 2022 17:26
Operator : JC/MD
Sample : N6001-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-12

Quant Time: Dec 14 05:40:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

