

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121322\
 Data File : VX033401.D
 Acq On : 13 Dec 2022 18:35
 Operator : JC/MD
 Sample : N6001-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-15

Quant Time: Dec 14 05:41:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	122127	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	195489	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	175290	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	82853	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	36519	46.611	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.220%
35) Dibromofluoromethane	5.391	113	40933	46.553	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.100%
50) Toluene-d8	8.647	98	83729	45.520	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.040%#
62) 4-Bromofluorobenzene	11.079	95	63456	45.119	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.240%
Target Compounds						
					Qvalue	
40) Benzene	6.038	78	6039	1.208	ug/l	100
67) Ethyl Benzene	10.195	91	1748	0.283	ug/l	100
68) m/p-Xylenes	10.305	106	1030	0.428	ug/l	99
69) o-Xylene	10.640	106	522	0.219	ug/l	74
84) 1,2,4-Trimethylbenzene	11.756	105	1113	0.223	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121322\
Data File : VX033401.D
Acq On : 13 Dec 2022 18:35
Operator : JC/MD
Sample : N6001-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-15

Quant Time: Dec 14 05:41:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

