

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121322\
 Data File : VX033403.D
 Acq On : 13 Dec 2022 19:21
 Operator : JC/MD
 Sample : N6001-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-17

Quant Time: Dec 14 05:41:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	121529	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	201786	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	177801	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	81333	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	37655	48.297	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.600%
35) Dibromofluoromethane	5.385	113	41571	45.804	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.600%
50) Toluene-d8	8.653	98	85078	44.810	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	89.620%#
62) 4-Bromofluorobenzene	11.079	95	64551	44.465	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.920%
Target Compounds						
					Qvalue	
19) Methyl tert-butyl Ether	3.117	73	3000	0.786	ug/l	98
30) Chloroform	5.098	83	3591	1.460	ug/l	89
47) Bromodichloromethane	7.823	83	984	0.502	ug/l #	74
60) Dibromochloromethane	9.524	129	362	0.234	ug/l	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121322\
Data File : VX033403.D
Acq On : 13 Dec 2022 19:21
Operator : JC/MD
Sample : N6001-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-17

Quant Time: Dec 14 05:41:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

