

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121322\
 Data File : VX033405.D
 Acq On : 13 Dec 2022 20:07
 Operator : JC/MD
 Sample : N6001-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-9

Quant Time: Dec 14 05:42:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	117134	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	189792	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	168498	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	74577	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	36464	48.524	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.040%
35) Dibromofluoromethane	5.385	113	40741	47.726	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.460%
50) Toluene-d8	8.647	98	83162	46.569	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.140%
62) 4-Bromofluorobenzene	11.079	95	61359	44.937	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121322\
Data File : VX033405.D
Acq On : 13 Dec 2022 20:07
Operator : JC/MD
Sample : N6001-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-9

Quant Time: Dec 14 05:42:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

