

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121323\
 Data File : VX039191.D
 Acq On : 13 Dec 2023 15:56
 Operator : JC/MD
 Sample : VX1213WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Dec 14 01:33:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	129687	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	237532	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	220174	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	105418	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	119956	50.491	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.980%
35) Dibromofluoromethane	5.385	113	85386	52.033	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.060%
50) Toluene-d8	8.647	98	321343	52.747	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.500%
62) 4-Bromofluorobenzene	11.079	95	127677	50.794	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.580%
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.166	85	35773	19.630	ug/l	100
3) Chloromethane	1.294	50	37508	18.566	ug/l	97
4) Vinyl Chloride	1.374	62	35713	19.007	ug/l	98
5) Bromomethane	1.605	94	23473	21.715	ug/l	98
6) Chloroethane	1.691	64	20322	17.948	ug/l	100
7) Trichlorofluoromethane	1.886	101	51306	18.648	ug/l	95
8) Diethyl Ether	2.130	74	17685	18.572	ug/l	96
9) 1,1,2-Trichlorotrifluo...	2.331	101	31596	18.819	ug/l	97
10) Methyl Iodide	2.453	142	34163	19.686	ug/l	95
11) Tert butyl alcohol	2.947	59	53258	89.270	ug/l #	93
12) 1,1-Dichloroethene	2.319	96	31629	18.965	ug/l	100
13) Acrolein	2.233	56	44468	93.431	ug/l	99
14) Allyl chloride	2.666	41	61595	18.232	ug/l	97
15) Acrylonitrile	3.062	53	110292	92.359	ug/l	98
16) Acetone	2.374	43	112061	63.624	ug/l	96
17) Carbon Disulfide	2.514	76	88864	18.208	ug/l	99
18) Methyl Acetate	2.697	43	97142	18.904	ug/l	100
19) Methyl tert-butyl Ether	3.111	73	111098	18.420	ug/l	97
20) Methylene Chloride	2.788	84	35493	17.921	ug/l	98
21) trans-1,2-Dichloroethene	3.093	96	33278	18.338	ug/l	94
22) Diisopropyl ether	3.757	45	112979	18.371	ug/l	97
23) Vinyl Acetate	3.721	43	432332	94.430	ug/l	100
24) 1,1-Dichloroethane	3.611	63	63981	18.405	ug/l	99
25) 2-Butanone	4.550	43	164180	80.083	ug/l	100
26) 2,2-Dichloropropane	4.471	77	55132	18.165	ug/l	100
27) cis-1,2-Dichloroethene	4.489	96	37263	18.604	ug/l	96
28) Bromochloromethane	4.898	49	28966	17.077	ug/l	93
29) Tetrahydrofuran	5.001	42	104568	92.361	ug/l	99
30) Chloroform	5.093	83	64093	18.449	ug/l	97
31) Cyclohexane	5.471	56	57958	18.888	ug/l	93
32) 1,1,1-Trichloroethane	5.385	97	55591	18.433	ug/l	99
36) 1,1-Dichloropropene	5.696	75	48868	18.871	ug/l	98
37) Ethyl Acetate	4.708	43	59857	18.710	ug/l	99
38) Carbon Tetrachloride	5.678	117	44816	18.135	ug/l	94
39) Methylcyclohexane	7.379	83	57619	18.524	ug/l	98
40) Benzene	6.038	78	137507	18.772	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121323\
 Data File : VX039191.D
 Acq On : 13 Dec 2023 15:56
 Operator : JC/MD
 Sample : VX1213WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Dec 14 01:33:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	32847	19.243	ug/l	97
42) 1,2-Dichloroethane	6.086	62	56373	19.012	ug/l	98
43) Isopropyl Acetate	6.336	43	94121	18.092	ug/l	100
44) Trichloroethene	7.123	130	33584	18.169	ug/l	94
45) 1,2-Dichloropropane	7.428	63	37180	18.912	ug/l	99
46) Dibromomethane	7.580	93	26506	19.180	ug/l	92
47) Bromodichloromethane	7.818	83	49644	18.788	ug/l	99
48) Methyl methacrylate	7.690	41	57949	19.053	ug/l	97
49) 1,4-Dioxane	7.659	88	15954	389.151	ug/l #	94
51) 4-Methyl-2-Pentanone	8.568	43	315832	95.282	ug/l	98
52) Toluene	8.714	92	84171	19.038	ug/l	95
53) t-1,3-Dichloropropene	8.976	75	54373	18.691	ug/l	96
54) cis-1,3-Dichloropropene	8.366	75	57809	18.469	ug/l	97
55) 1,1,2-Trichloroethane	9.153	97	34449	18.664	ug/l	94
56) Ethyl methacrylate	9.116	69	43640	19.368	ug/l	98
57) 1,3-Dichloropropane	9.305	76	60929	19.094	ug/l	98
58) 2-Chloroethyl Vinyl ether	8.238	63	135035	102.061	ug/l	94
59) 2-Hexanone	9.427	43	252131	88.136	ug/l	99
60) Dibromochloromethane	9.519	129	35251	18.884	ug/l	98
61) 1,2-Dibromoethane	9.610	107	36641	18.596	ug/l	98
64) Tetrachloroethene	9.275	164	26536	18.444	ug/l	96
65) Chlorobenzene	10.079	112	86972	18.098	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.159	131	30713	18.093	ug/l	98
67) Ethyl Benzene	10.195	91	168610	18.848	ug/l	98
68) m/p-Xylenes	10.299	106	119872	37.374	ug/l	94
69) o-Xylene	10.640	106	57513	18.546	ug/l	92
70) Styrene	10.653	104	84736	18.691	ug/l	95
71) Bromoform	10.799	173	22900	18.149	ug/l #	100
73) Isopropylbenzene	10.963	105	157805	19.402	ug/l	98
74) N-amyl acetate	10.842	43	77343	18.795	ug/l	97
75) 1,1,2,2-Tetrachloroethane	11.207	83	58012	19.173	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	65619	24.060	ug/l	91
77) Bromobenzene	11.195	156	33605	18.684	ug/l	88
78) n-propylbenzene	11.305	91	192449	19.204	ug/l	97
79) 2-Chlorotoluene	11.360	91	112904	18.626	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	134631	19.431	ug/l	96
81) trans-1,4-Dichloro-2-b...	11.018	75	16777	17.353	ug/l	99
82) 4-Chlorotoluene	11.451	91	137617	19.330	ug/l	96
83) tert-Butylbenzene	11.713	119	125227	18.991	ug/l	95
84) 1,2,4-Trimethylbenzene	11.750	105	133634	18.807	ug/l	99
85) sec-Butylbenzene	11.890	105	159908	19.057	ug/l	97
86) p-Isopropyltoluene	12.006	119	131651	18.695	ug/l	97
87) 1,3-Dichlorobenzene	11.969	146	66707	18.879	ug/l	99
88) 1,4-Dichlorobenzene	12.043	146	67274	18.286	ug/l	97
89) n-Butylbenzene	12.329	91	133509	19.286	ug/l	96
90) Hexachloroethane	12.536	117	23385	18.384	ug/l	91
91) 1,2-Dichlorobenzene	12.335	146	62267	18.417	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.939	75	14375	18.038	ug/l	95
93) 1,2,4-Trichlorobenzene	13.585	180	41062	18.548	ug/l	99
94) Hexachlorobutadiene	13.725	225	15114	18.116	ug/l	97
95) Naphthalene	13.774	128	147058	19.074	ug/l	99
96) 1,2,3-Trichlorobenzene	13.957	180	39175	18.330	ug/l	97


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121323\
Data File : VX039191.D
Acq On : 13 Dec 2023 15:56
Operator : JC/MD
Sample : VX1213WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Dec 14 01:33:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:168 Resp: 129687
Ion Ratio Lower Upper
168 100
99 69.0 56.6 85.0



#2
Dichlorodifluoromethane
Concen: 19.630 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56

Tgt Ion: 85 Resp: 35773
Ion Ratio Lower Upper
85 100
87 33.6 16.7 50.1









#3
Chloromethane
Concen: 18.566 ug/l
RT: 1.294 min Scan# 34
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 50 Resp: 37508
Ion Ratio Lower Upper
50 100
52 33.7 25.5 38.3



#4
Vinyl Chloride
Concen: 19.007 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56

Tgt Ion: 62 Resp: 35713
Ion Ratio Lower Upper
62 100
64 30.9 25.4 38.0





#5

Bromomethane

Concen: 21.715 ug/l

RT: 1.605 min Scan# 84

Delta R.T. 0.006 min

Lab File: VX039191.D

Acq: 13 Dec 2023 15:56

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 94 Resp: 23473

Ion Ratio Lower Upper

94 100

96 91.5 74.6 111.8



#6

Chloroethane

Concen: 17.948 ug/l

RT: 1.691 min Scan# 99

Delta R.T. 0.006 min

Lab File: VX039191.D




Acq: 13 Dec 2023 15:56


Tgt Ion: 64 Resp: 20322

Ion Ratio Lower Upper

64 100


66 31.6 25.1 37.7








#7
 Trichlorofluoromethane
 Concen: 18.648 ug/l
 RT: 1.886 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: VX039191.D
 Acq: 13 Dec 2023 15:56

Instrument :
 MSVOA_X
 ClientSampleId :







Tgt Ion:101 Resp: 51306
 Ion Ratio Lower Upper
 101 100
 103 60.8 51.4 77.2



#8
 Diethyl Ether
 Concen: 18.572 ug/l
 RT: 2.130 min Scan# 171
 Delta R.T. 0.000 min
 Lab File: VX039191.D
 Acq: 13 Dec 2023 15:56

Tgt Ion: 74 Resp: 17685
 Ion Ratio Lower Upper
 74 100
 45 102.7 53.3 159.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 18.819 ug/l
RT: 2.331 min Scan# 204
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:101 Resp: 31596
Ion Ratio Lower Upper
101 100
85 46.6 37.0 55.6
151 69.9 58.6 87.8



#10
Methyl Iodide
Concen: 19.686 ug/l
RT: 2.453 min Scan# 224
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56

Tgt Ion:142 Resp: 34163
Ion Ratio Lower Upper
142 100
127 48.1 41.6 62.4
141 13.8 11.6 17.4





#11

Tert butyl alcohol

Concen: 89.270 ug/l

RT: 2.947 min Scan# 306

Delta R.T. -0.006 min

Lab File: VX039191.D

Acq: 13 Dec 2023 15:56

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 59 Resp: 53258

Ion Ratio Lower Upper

59 100

57 9.8 6.0 9.0#



#12

1,1-Dichloroethene

Concen: 18.965 ug/l

RT: 2.319 min Scan# 202

Delta R.T. 0.000 min

Lab File: VX039191.D

Acq: 13 Dec 2023 15:56




Tgt Ion: 96 Resp: 31629


Ion Ratio Lower Upper

96 100

61 181.1 144.6 216.8

98 63.7 50.4 75.6





#13

Acrolein

Concen: 93.431 ug/l

RT: 2.233 min Scan# 188

Delta R.T. 0.000 min

Lab File: VX039191.D

Acq: 13 Dec 2023 15:56

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 56 Resp: 44468

Ion Ratio Lower Upper

56 100

55 70.8 57.1 85.7



#14

Allyl chloride

Concen: 18.232 ug/l

RT: 2.666 min Scan# 259

Delta R.T. 0.000 min

Lab File: VX039191.D

Acq: 13 Dec 2023 15:56




Tgt Ion: 41 Resp: 61595


Ion Ratio Lower Upper

41 100

39 67.2 55.8 83.6


76 31.4 24.3 36.5








#15
Acrylonitrile
Concen: 92.359 ug/l
RT: 3.062 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56

Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion: 53 Resp: 110292




Ion	Ratio	Lower	Upper
53	100		
52	84.2	66.0	99.0
51	36.4	29.8	44.6




#16
Acetone
Concen: 63.624 ug/l
RT: 2.374 min Scan# 211
Delta R.T. 0.001 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56

Tgt Ion: 43 Resp: 112061


Ion	Ratio	Lower	Upper
43	100		
58	29.3	21.9	32.9








#17
Carbon Disulfide
Concen: 18.208 ug/l
RT: 2.514 min Scan# 21
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion: 76 Resp: 88864
Ion Ratio Lower Upper
76 100
78 9.2 7.0 10.6




#18
Methyl Acetate
Concen: 18.904 ug/l
RT: 2.697 min Scan# 264
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56



Tgt Ion: 43 Resp: 97142
Ion Ratio Lower Upper
43 100
74 22.5 17.8 26.8









#19
Methyl tert-butyl Ether
Concen: 18.420 ug/l
RT: 3.111 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 73 Resp: 111098
Ion Ratio Lower Upper
73 100
57 23.3 17.3 25.9



#20
Methylene Chloride
Concen: 17.921 ug/l
RT: 2.788 min Scan# 279
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56

Tgt Ion: 84 Resp: 35493
Ion Ratio Lower Upper
84 100
49 131.0 103.6 155.4
51 41.2 33.8 50.6
86 63.7 49.4 74.0









#21
trans-1,2-Dichloroethene
Concen: 18.338 ug/l
RT: 3.093 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 96 Resp: 33278
Ion Ratio Lower Upper
96 100
61 157.4 119.2 178.8
98 63.4 48.7 73.1



#22
Diisopropyl ether
Concen: 18.371 ug/l
RT: 3.757 min Scan# 438
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56

Tgt Ion: 45 Resp: 112979
Ion Ratio Lower Upper
45 100
43 83.9 69.7 104.5
87 27.4 20.8 31.2
59 11.4 9.0 13.4









#23
 Vinyl Acetate
 Concen: 94.430 ug/l
 RT: 3.721 min Scan# 41
 Delta R.T. 0.000 min
 Lab File: VX039191.D
 Acq: 13 Dec 2023 15:56




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 43 Resp: 432332
 Ion Ratio Lower Upper
 43 100
 86 9.6 7.8 11.8



#24
 1,1-Dichloroethane
 Concen: 18.405 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. 0.006 min
 Lab File: VX039191.D
 Acq: 13 Dec 2023 15:56

Tgt Ion: 63 Resp: 63981
 Ion Ratio Lower Upper
 63 100
 98 6.4 3.3 9.8
 100 3.6 2.1 6.5









#25
2-Butanone
Concen: 80.083 ug/l
RT: 4.550 min Scan# 568
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 164180
Ion Ratio Lower Upper
43 100
72 24.3 19.4 29.2



#26
2,2-Dichloropropane
Concen: 18.165 ug/l
RT: 4.471 min Scan# 555
Delta R.T. -0.006 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56


Tgt Ion: 77 Resp: 55132
Ion Ratio Lower Upper
77 100
97 21.5 10.8 32.3








#27
 cis-1,2-Dichloroethene
 Concen: 18.604 ug/l
 RT: 4.489 min Scan# 511
 Delta R.T. 0.006 min
 Lab File: VX039191.D
 Acq: 13 Dec 2023 15:56

Instrument :
 MSVOA_X
 ClientSampleId :






Tgt Ion: 96 Resp: 37263
 Ion Ratio Lower Upper
 96 100
 61 163.0 0.0 312.2
 98 66.0 0.0 129.0



#28
 Bromochloromethane
 Concen: 17.077 ug/l
 RT: 4.898 min Scan# 625
 Delta R.T. 0.000 min
 Lab File: VX039191.D
 Acq: 13 Dec 2023 15:56

Tgt Ion: 49 Resp: 28966
 Ion Ratio Lower Upper
 49 100
 129 0.7 0.0 3.8
 130 65.8 57.5 86.3





#29

Tetrahydrofuran

Concen: 92.361 ug/l

RT: 5.001 min Scan# 641

Delta R.T. 0.006 min

Lab File: VX039191.D

Acq: 13 Dec 2023 15:56

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 42 Resp: 104568

Ion Ratio Lower Upper

42 100

72 42.2 34.2 51.4

71 40.5 32.2 48.2



#30

Chloroform

Concen: 18.449 ug/l

RT: 5.093 min Scan# 657

Delta R.T. 0.000 min

Lab File: VX039191.D




Acq: 13 Dec 2023 15:56


Tgt Ion: 83 Resp: 64093

Ion Ratio Lower Upper

83 100


85 65.4 50.3 75.5








#31
Cyclohexane
Concen: 18.888 ug/l
RT: 5.471 min Scan# 719
Delta R.T. -0.006 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion: 56 Resp: 57958
Ion Ratio Lower Upper
56 100
69 28.6 23.4 35.2
84 74.0 65.9 98.9




#32
1,1,1-Trichloroethane
Concen: 18.433 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.006 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56



Tgt Ion: 97 Resp: 55591
Ion Ratio Lower Upper
97 100
99 63.9 51.1 76.7
61 47.6 36.4 54.6





#33

1,2-Dichloroethane-d4

Concen: 50.491 ug/l

RT: 5.952 min Scan# 798

Delta R.T. 0.000 min

Lab File: VX039191.D

Acq: 13 Dec 2023 15:56

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 65 Resp: 119956

Ion Ratio Lower Upper

65 100

67 52.6 0.0 100.2



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.006 min

Lab File: VX039191.D

Acq: 13 Dec 2023 15:56




Tgt Ion: 114 Resp: 237532


Ion Ratio Lower Upper

114 100

63 24.6 0.0 45.0


88 18.3 0.0 36.6








#35
Dibromofluoromethane
Concen: 52.033 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56

Instrument :
MSVOA_X
ClientSampleId :






Tgt Ion:113 Resp: 85386
Ion Ratio Lower Upper
113 100
111 102.3 82.9 124.3
192 16.1 13.3 19.9



#36
1,1-Dichloropropene
Concen: 18.871 ug/l
RT: 5.696 min Scan# 756
Delta R.T. 0.006 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56

Tgt Ion: 75 Resp: 48868
Ion Ratio Lower Upper
75 100
110 32.9 17.0 50.9
77 29.8 24.7 37.1





#37

Ethyl Acetate

Concen: 18.710 ug/l

RT: 4.708 min Scan# 594

Delta R.T. 0.000 min


Lab File: VX039191.D

Acq: 13 Dec 2023 15:56




Instrument :

MSVOA_X

ClientSampleId :



Tgt Ion: 43	Resp: 59857
Ion Ratio	Lower Upper
43	100
61	13.8 10.6 16.0
70	10.1 8.3 12.5



#38

Carbon Tetrachloride

Concen: 18.135 ug/l




RT: 5.678 min Scan# 753


Delta R.T. 0.000 min

Lab File: VX039191.D

Acq: 13 Dec 2023 15:56

Tgt Ion: 117	Resp: 44816
Ion Ratio	Lower Upper
117	100
119	91.2 78.2 117.2
121	27.6 24.4 36.6









#39
Methylcyclohexane
Concen: 18.524 ug/l
RT: 7.379 min Scan# 1032
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 57619
Ion Ratio Lower Upper
83 100
55 82.3 64.6 97.0
98 44.6 36.7 55.1



#40
Benzene
Concen: 18.772 ug/l
RT: 6.038 min Scan# 812
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56


Tgt Ion: 78 Resp: 137507
Ion Ratio Lower Upper
78 100
77 25.9 19.3 28.9








#41
Methacrylonitrile
Concen: 19.243 ug/l
RT: 4.916 min Scan# 61
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56

Instrument :
MSVOA_X
ClientSampleId :




Tgt Ion: 41 Resp: 32847

Ion	Ratio	Lower	Upper
41	100		
39	54.7	45.2	67.8
67	68.9	52.6	78.8
52	30.6	22.7	34.1






#42
1,2-Dichloroethane
Concen: 19.012 ug/l
RT: 6.086 min Scan# 820
Delta R.T. 0.006 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56



Tgt Ion: 62 Resp: 56373

Ion	Ratio	Lower	Upper
62	100		
98	7.2	0.0	16.0









#43
Isopropyl Acetate
Concen: 18.092 ug/l
RT: 6.336 min Scan# 861
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion: 43 Resp: 94121
Ion Ratio Lower Upper
43 100
61 19.0 15.2 22.8
87 12.1 9.8 14.6



#44
Trichloroethene
Concen: 18.169 ug/l
RT: 7.123 min Scan# 990
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56

Tgt Ion:130 Resp: 33584
Ion Ratio Lower Upper
130 100
95 108.5 0.0 204.4









#45
1,2-Dichloropropane
Concen: 18.912 ug/l
RT: 7.428 min Scan# 1040
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 63 Resp: 37180
Ion Ratio Lower Upper
63 100
65 30.1 24.3 36.5



#46
Dibromomethane
Concen: 19.180 ug/l
RT: 7.580 min Scan# 1065
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56

Tgt Ion: 93 Resp: 26506
Ion Ratio Lower Upper
93 100
95 79.4 66.6 100.0
174 76.7 69.4 104.0









#47
Bromodichloromethane
Concen: 18.788 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 49644
Ion Ratio Lower Upper
83 100
85 61.7 49.8 74.6
127 7.5 7.1 10.7



#48
Methyl methacrylate
Concen: 19.053 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56

Tgt Ion: 41 Resp: 57949
Ion Ratio Lower Upper
41 100
69 77.1 59.2 88.8
39 61.7 47.7 71.5





#49

1,4-Dioxane

Concen: 389.151 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. 0.006 min

Lab File: VX039191.D

Acq: 13 Dec 2023 15:56

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 88 Resp: 15954

Ion Ratio Lower Upper

88 100

43 40.8 0.0 0.0#

58 80.0 60.1 90.1



#50

Toluene-d8

Concen: 52.747 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. 0.000 min

Lab File: VX039191.D




Acq: 13 Dec 2023 15:56


Tgt Ion: 98 Resp: 321343

Ion Ratio Lower Upper

98 100

100 63.3 53.1 79.7









#51
4-Methyl-2-Pentanone
Concen: 95.282 ug/l
RT: 8.568 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 315832
Ion Ratio Lower Upper
43 100
58 37.1 30.4 45.6



#52
Toluene
Concen: 19.038 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.006 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56

Tgt Ion: 92 Resp: 84171
Ion Ratio Lower Upper
92 100
91 174.7 134.2 201.2





#53

t-1,3-Dichloropropene

Concen: 18.691 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX039191.D

Acq: 13 Dec 2023 15:56

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 75 Resp: 54373

Ion Ratio Lower Upper

75 100

77 29.7 25.7 38.5



#54

cis-1,3-Dichloropropene

Concen: 18.469 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX039191.D

Acq: 13 Dec 2023 15:56




Tgt Ion: 75 Resp: 57809


Ion Ratio Lower Upper

75 100

77 32.2 25.3 37.9


39 61.2 46.7 70.1








#55
1,1,2-Trichloroethane
Concen: 18.664 ug/l
RT: 9.153 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion: 97	Resp: 34449
Ion Ratio	Lower Upper
97 100	
83 89.2	65.8 98.8
85 57.3	41.6 62.4
99 62.0	47.8 71.8




#56
Ethyl methacrylate
Concen: 19.368 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56



Tgt Ion: 69	Resp: 43640
Ion Ratio	Lower Upper
69 100	
41 70.5	56.8 85.2
39 42.8	32.0 48.0





#57

1,3-Dichloropropane

Concen: 19.094 ug/l

RT: 9.305 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX039191.D

Acq: 13 Dec 2023 15:56

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 76 Resp: 60929

Ion Ratio Lower Upper

76 100

78 31.5 26.1 39.1



#58

2-Chloroethyl Vinyl ether

Concen: 102.061 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. 0.000 min

Lab File: VX039191.D




Acq: 13 Dec 2023 15:56


Tgt Ion: 63 Resp: 135035

Ion Ratio Lower Upper

63 100

106 25.0 22.5 33.7









#59
2-Hexanone
Concen: 88.136 ug/l
RT: 9.427 min Scan# 1368
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 252131
Ion Ratio Lower Upper
43 100
58 51.7 26.2 78.5



#60
Dibromochloromethane
Concen: 18.884 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56

Tgt Ion:129 Resp: 35251
Ion Ratio Lower Upper
129 100
127 75.8 38.8 116.4









#61
1,2-Dibromoethane
Concen: 18.596 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:107 Resp: 36641
Ion Ratio Lower Upper
107 100
109 96.0 75.2 112.8



#62
4-Bromofluorobenzene
Concen: 50.794 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56


Tgt Ion: 95 Resp: 127677
Ion Ratio Lower Upper
95 100
174 62.3 0.0 135.6
176 60.4 0.0 131.0








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:117 Resp: 220174
Ion Ratio Lower Upper
117 100
82 60.1 46.2 69.4
119 32.0 25.5 38.3



#64
Tetrachloroethene
Concen: 18.444 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56

Tgt Ion:164 Resp: 26536
Ion Ratio Lower Upper
164 100
166 130.2 102.7 154.1
129 105.1 80.4 120.6
131 101.9 76.5 114.7









#65
Chlorobenzene
Concen: 18.098 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:112 Resp: 86972
Ion Ratio Lower Upper
112 100
114 33.1 25.7 38.5



#66
1,1,1,2-Tetrachloroethane
Concen: 18.093 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56

Tgt Ion:131 Resp: 30713
Ion Ratio Lower Upper
131 100
133 94.3 47.1 141.4
119 67.5 32.2 96.6









#67
Ethyl Benzene
Concen: 18.848 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 168610
Ion Ratio Lower Upper
91 100
106 29.9 24.9 37.3



#68
m/p-Xylenes
Concen: 37.374 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56

Tgt Ion: 106 Resp: 119872
Ion Ratio Lower Upper
106 100
91 215.8 165.5 248.3





#69

o-Xylene

Concen: 18.546 ug/l

RT: 10.640 min Scan# 1567

Delta R.T. 0.000 min

Lab File: VX039191.D

Acq: 13 Dec 2023 15:56

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion:106 Resp: 57513

Ion Ratio Lower Upper

106 100

91 231.0 108.9 326.8



#70

Styrene

Concen: 18.691 ug/l

RT: 10.653 min Scan# 1569

Delta R.T. 0.000 min

Lab File: VX039191.D

Acq: 13 Dec 2023 15:56




Tgt Ion:104 Resp: 84736


Ion Ratio Lower Upper

104 100

78 59.5 43.6 65.4


103 56.5 43.6 65.4








#71
Bromoform
Concen: 18.149 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:173 Resp: 22900
Ion Ratio Lower Upper
173 100
175 48.6 24.1 72.4
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56

Tgt Ion:152 Resp: 105418
Ion Ratio Lower Upper
152 100
115 79.4 43.3 129.9
150 164.1 0.0 345.0





#73

Isopropylbenzene

Concen: 19.402 ug/l

RT: 10.963 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX039191.D

Acq: 13 Dec 2023 15:56

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion:105 Resp: 157805

Ion Ratio Lower Upper

105 100

120 24.8 12.9 38.6



#74

N-ethyl acetate

Concen: 18.795 ug/l

RT: 10.842 min Scan# 1600

Delta R.T. 0.000 min

Lab File: VX039191.D

Acq: 13 Dec 2023 15:56

Tgt Ion: 43 Resp: 77343




Ion Ratio Lower Upper


43 100

70 38.3 32.9 49.3

55 23.2 19.3 28.9

61 22.2 18.8 28.2





#75

1,1,2,2-Tetrachloroethane

Concen: 19.173 ug/l

RT: 11.207 min Scan# 1

Delta R.T. -0.006 min

Lab File: VX039191.D





Acq: 13 Dec 2023 15:56

Instrument :

MSVOA_X

ClientSampleId :

Tgt Ion	Ratio	Lower	Upper
83	100		
131	9.4	5.3	15.8
85	64.3	32.0	96.2



#76

1,2,3-Trichloropropane

Concen: 24.060 ug/l




RT: 11.238 min Scan# 1665


Delta R.T. 0.000 min

Lab File: VX039191.D

Acq: 13 Dec 2023 15:56

Tgt Ion	Ratio	Lower	Upper
75	100		
77	33.0	19.1	57.5





#77

Bromobenzene

Concen: 18.684 ug/l

RT: 11.195 min Scan# 1658

Delta R.T. 0.000 min

Lab File: VX039191.D

Acq: 13 Dec 2023 15:56

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:156 Resp: 33605

Ion Ratio Lower Upper

156 100

77 202.1 88.4 265.3

158 98.2 48.0 144.0



#78

n-propylbenzene

Concen: 19.204 ug/l

RT: 11.305 min Scan# 1676

Delta R.T. 0.000 min

Lab File: VX039191.D




Acq: 13 Dec 2023 15:56


Tgt Ion: 91 Resp: 192449

Ion Ratio Lower Upper

91 100

120 21.3 11.5 34.4









#79
2-Chlorotoluene
Concen: 18.626 ug/l
RT: 11.360 min Scan# 1685
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 112904
Ion Ratio Lower Upper
91 100
126 32.0 16.3 48.8



#80
1,3,5-Trimethylbenzene
Concen: 19.431 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56

Tgt Ion:105 Resp: 134631
Ion Ratio Lower Upper
105 100
120 45.4 24.1 72.3









#81
trans-1,4-Dichloro-2-butene
Concen: 17.353 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion: 75 Resp: 16777
Ion Ratio Lower Upper
75 100
53 111.3 90.5 135.7
89 46.3 36.9 55.3



#82
4-Chlorotoluene
Concen: 19.330 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX039191.D
Acq: 13 Dec 2023 15:56

Tgt Ion: 91 Resp: 137617
Ion Ratio Lower Upper
91 100
126 26.3 14.2 42.6





#83

tert-Butylbenzene

Concen: 18.991 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. 0.000 min

Lab File: VX039191.D

Acq: 13 Dec 2023 15:56

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:119 Resp: 125227

Ion Ratio Lower Upper

119 100

91 67.2 31.2 93.6

134 22.0 11.7 35.0



#84

1,2,4-Trimethylbenzene

Concen: 18.807 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. 0.000 min

Lab File: VX039191.D




Acq: 13 Dec 2023 15:56


Tgt Ion:105 Resp: 133634

Ion Ratio Lower Upper

105 100

120 42.9 21.2 63.6









#85
 sec-Butylbenzene
 Concen: 19.057 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX039191.D
 Acq: 13 Dec 2023 15:56




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion:105 Resp: 159908
 Ion Ratio Lower Upper
 105 100
 134 18.1 9.8 29.4



#86
 p-Isopropyltoluene
 Concen: 18.695 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. 0.000 min
 Lab File: VX039191.D
 Acq: 13 Dec 2023 15:56


Tgt Ion:119 Resp: 131651
 Ion Ratio Lower Upper
 119 100
 134 25.2 12.8 38.3
 91 28.5 13.1 39.3








#87
 1,3-Dichlorobenzene
 Concen: 18.879 ug/l
 RT: 11.969 min Scan# 1785
 Delta R.T. 0.000 min
 Lab File: VX039191.D
 Acq: 13 Dec 2023 15:56

Instrument :
 MSVOA_X
 ClientSampleId :







Tgt Ion:146 Resp: 66707
 Ion Ratio Lower Upper
 146 100
 111 45.5 22.1 66.1
 148 64.1 31.8 95.3



#88
 1,4-Dichlorobenzene
 Concen: 18.286 ug/l
 RT: 12.043 min Scan# 1797
 Delta R.T. 0.006 min
 Lab File: VX039191.D
 Acq: 13 Dec 2023 15:56

Tgt Ion:146 Resp: 67274
 Ion Ratio Lower Upper
 146 100
 111 47.4 21.3 63.7
 148 63.5 31.8 95.4





#89

n-Butylbenzene

Concen: 19.286 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. 0.000 min

Lab File: VX039191.D

Acq: 13 Dec 2023 15:56

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 91 Resp: 133509

Ion Ratio Lower Upper

91 100

92 51.9 27.4 82.0

134 22.9 12.6 37.8



#90

Hexachloroethane

Concen: 18.384 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX039191.D




Acq: 13 Dec 2023 15:56


Tgt Ion: 117 Resp: 23385

Ion Ratio Lower Upper

117 100

201 58.8 32.9 98.6





#91

1,2-Dichlorobenzene

Concen: 18.417 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.000 min

Lab File: VX039191.D

Acq: 13 Dec 2023 15:56

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:146 Resp: 62267

Ion Ratio Lower Upper

146 100

111 47.5 22.4 67.0

148 63.5 31.6 95.0



#92

1,2-Dibromo-3-Chloropropane

Concen: 18.038 ug/l

RT: 12.939 min Scan# 1944

Delta R.T. 0.000 min

Lab File: VX039191.D

Acq: 13 Dec 2023 15:56




Tgt Ion: 75 Resp: 14375


Ion Ratio Lower Upper

75 100

155 67.2 36.1 108.5

157 86.8 44.9 134.7





#93

1,2,4-Trichlorobenzene

Concen: 18.548 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. 0.000 min

Lab File: VX039191.D

Acq: 13 Dec 2023 15:56

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:180 Resp: 41062

Ion Ratio Lower Upper

180 100

182 95.4 48.3 144.8

145 36.3 18.0 54.0



#94

Hexachlorobutadiene

Concen: 18.116 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. 0.000 min

Lab File: VX039191.D

Acq: 13 Dec 2023 15:56




Tgt Ion:225 Resp: 15114


Ion Ratio Lower Upper

225 100

223 63.0 31.1 93.2

227 61.7 32.5 97.5





#95

Naphthalene

Concen: 19.074 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. 0.000 min

Lab File: VX039191.D

Acq: 13 Dec 2023 15:56

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:128 Resp: 147058

Ion Ratio Lower Upper

128 100

127 12.7 10.3 15.5

129 10.7 8.8 13.2



#96

1,2,3-Trichlorobenzene

Concen: 18.330 ug/l

RT: 13.957 min Scan# 2111

Delta R.T. 0.000 min

Lab File: VX039191.D

Acq: 13 Dec 2023 15:56

Tgt Ion:180 Resp: 39175

Ion Ratio Lower Upper

180 100

182 95.3 46.9 140.6

145 39.9 17.8 53.5

