

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121323\
 Data File : VX039186.D
 Acq On : 13 Dec 2023 12:29
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Dec 14 01:30:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	109539	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	222018	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	227877	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	113871	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	96544	48.111	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.220%
35) Dibromofluoromethane	5.385	113	72930	47.548	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.100%
50) Toluene-d8	8.647	98	280468	49.254	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.500%
62) 4-Bromofluorobenzene	11.079	95	128977	54.897	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121323\
Data File : VX039186.D
Acq On : 13 Dec 2023 12:29
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Dec 14 01:30:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

