

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121323\
 Data File : VX039189.D
 Acq On : 13 Dec 2023 15:11
 Operator : JC/MD
 Sample : VX1213WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1213WBL01

Quant Time: Dec 14 01:32:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	112881	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	231638	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	237362	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	117591	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	100554	48.626	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.260%
35) Dibromofluoromethane	5.385	113	76755	47.964	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.920%
50) Toluene-d8	8.647	98	291276	49.028	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.060%
62) 4-Bromofluorobenzene	11.079	95	133027	54.269	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121323\
Data File : VX039189.D
Acq On : 13 Dec 2023 15:11
Operator : JC/MD
Sample : VX1213WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1213WBL01

Quant Time: Dec 14 01:32:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

