

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121324\
 Data File : VX044269.D
 Acq On : 13 Dec 2024 10:12
 Operator : JC/MD
 Sample : VX1213WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK736

Quant Time: Dec 13 22:47:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML120524WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Dec 07 03:56:15 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	195313	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.049	117	171491	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.018	152	69923	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	77266	43.229	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.460%
7) Chloroethane-d5	1.648	69	61600	44.491	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	88.980%
11) 1,1-Dichloroethene-d2	2.294	65	32517	42.540	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.080%
21) 2-Butanone-d5	4.458	46	111144	101.566	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	101.570%
24) Chloroform-d	5.050	84	150741	44.419	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.840%
26) 1,2-Dichloroethane-d4	5.952	65	112390	46.831	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.660%
32) Benzene-d6	5.964	84	280521	47.341	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.680%
36) 1,2-Dichloropropane-d6	7.306	67	90110	48.975	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.940%
41) Toluene-d8	8.641	98	241847	44.141	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.280%
43) trans-1,3-Dichloroprop...	8.945	79	37618	42.545	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.080%
47) 2-Hexanone-d5	9.384	63	67777	94.475	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	94.480%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	121774	49.359	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.720%
66) 1,2-Dichlorobenzene-d4	12.317	152	78967	48.471	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.940%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121324\
Data File : VX044269.D
Acq On : 13 Dec 2024 10:12
Operator : JC/MD
Sample : VX1213WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK736

Quant Time: Dec 13 22:47:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM120524WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Dec 07 03:56:15 2024
Response via : Initial Calibration

