

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX121418\
 Data File : VX006516.D
 Acq On : 14 Dec 2018 14:18
 Operator : JC/MD
 Sample : J6403-65 4X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-3-OILY-WATER

Quant Time: Dec 17 12:37:23 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	656304	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1041978	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	957539	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	477313	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	355959	53.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.26%	
35) Dibromofluoromethane	5.50	113	310248	45.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.98%	
50) Toluene-d8	8.71	98	1228971	50.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.16%	
62) 4-Bromofluorobenzene	11.13	95	461894	49.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.90%	
Target Compounds						
67) Ethyl Benzene	10.25	91	19299	0.571	ug/l	94
68) m/p-Xylenes	10.36	106	3875	0.286	ug/l	93
69) o-Xylene	10.70	106	7890	0.590	ug/l	99
73) Isopropylbenzene	11.02	105	11732	0.356	ug/l	95
80) 1,3,5-Trimethylbenzene	11.51	105	33198	1.210	ug/l	99
84) 1,2,4-Trimethylbenzene	11.80	105	108929	3.846	ug/l	99
86) p-Isopropyltoluene	12.06	119	13411	0.447	ug/l	85
89) n-Butylbenzene	12.37	91	15832	0.619	ug/l #	1
95) Naphthalene	13.83	128	2331065	63.585	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX121418\
Data File : VX006516.D
Acq On : 14 Dec 2018 14:18
Operator : JC/MD
Sample : J6403-65 4X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-3-OILY-WATER

Quant Time: Dec 17 12:37:23 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

