

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX121418\
 Data File : VX006527.D
 Acq On : 14 Dec 2018 19:38
 Operator : JC/MD
 Sample : J6396-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-14C_R1

Quant Time: Dec 15 05:30:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	653147	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1046260	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	942896	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	444440	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	357200	53.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.14%	
35) Dibromofluoromethane	5.51	113	309679	45.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.42%	
50) Toluene-d8	8.71	98	1232982	50.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.08%	
62) 4-Bromofluorobenzene	11.13	95	449118	48.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.74%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX121418\
Data File : VX006527.D
Acq On : 14 Dec 2018 19:38
Operator : JC/MD
Sample : J6396-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-14C_R1

Quant Time: Dec 15 05:30:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

