

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX121418\
 Data File : VX006530.D
 Acq On : 14 Dec 2018 20:58
 Operator : JC/MD
 Sample : J6396-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP3-2018-12-12

Quant Time: Dec 15 05:32:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	653844	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1050400	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	947222	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	453243	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	359943	53.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.86%	
35) Dibromofluoromethane	5.51	113	311558	45.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.62%	
50) Toluene-d8	8.71	98	1244594	50.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.62%	
62) 4-Bromofluorobenzene	11.13	95	455062	48.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.64%	
Target Compounds						
16) Acetone	2.45	43	5319	1.698	ug/l #	Qvalue 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX121418\
Data File : VX006530.D
Acq On : 14 Dec 2018 20:58
Operator : JC/MD
Sample : J6396-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP3-2018-12-12

Quant Time: Dec 15 05:32:55 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

