

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX121418\
 Data File : VX006552.D
 Acq On : 15 Dec 2018 08:47
 Operator : JC/MD
 Sample : J6398-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 C-MW-09-121118

Quant Time: Dec 15 23:40:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	657214	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1036137	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	954229	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	445930	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	356878	53.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.38%	
35) Dibromofluoromethane	5.50	113	309362	46.11	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.22%	
50) Toluene-d8	8.71	98	1236049	50.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.30%	
62) 4-Bromofluorobenzene	11.14	95	451594	49.11	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.22%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX121418\
Data File : VX006552.D
Acq On : 15 Dec 2018 08:47
Operator : JC/MD
Sample : J6398-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
C-MW-09-121118

Quant Time: Dec 15 23:40:07 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

