

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX121418\
 Data File : VX006534.D
 Acq On : 14 Dec 2018 22:45
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 15 05:41:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	69	0.00
2 T	Dichlorodifluoromethane	0.490	0.476	2.9	62	0.00
3 P	Chloromethane	0.503	0.511	-1.6	69	0.00
4 C	Vinyl Chloride	0.541	0.570	-5.4#	71	0.00
5 T	Bromomethane	0.375	0.539	-43.7#	106	0.00
6 T	Chloroethane	0.332	0.362	-9.0	73	0.00
7 T	Trichlorofluoromethane	0.831	0.892	-7.3	74	0.00
8 T	Diethyl Ether	0.315	0.346	-9.8	78	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.473	0.515	-8.9	77	0.00
10 T	Methyl Iodide	0.712	0.838	-17.7	80	0.00
11 T	Tert butyl alcohol	0.140	0.138	1.4	70	0.00
12 CM	1,1-Dichloroethene	0.452	0.491	-8.6#	77	0.00
13 T	Acrolein	0.076	0.078	-2.6	73	0.00
14 T	Allyl chloride	0.894	0.821	8.2	64	0.00
15 T	Acrylonitrile	0.291	0.306	-5.2	73	0.00
16 T	Acetone	0.240	0.248	-3.3	72	0.00
17 T	Carbon Disulfide	1.394	1.248	10.5	62	0.00
18 T	Methyl Acetate	0.801	0.835	-4.2	73	0.00
19 T	Methyl tert-butyl Ether	1.614	1.680	-4.1	73	0.00
20 T	Methylene Chloride	0.510	0.557	-9.2	80	0.00
21 T	trans-1,2-Dichloroethene	0.482	0.527	-9.3	77	0.00
22 T	Diisopropyl ether	1.396	1.539	-10.2	77	0.00
23 T	Vinyl Acetate	1.191	1.245	-4.5	72	0.00
24 P	1,1-Dichloroethane	0.798	0.881	-10.4	79	0.00
25 T	2-Butanone	0.342	0.373	-9.1	76	0.00
26 T	2,2-Dichloropropane	0.736	0.624	15.2	60	0.00
27 T	cis-1,2-Dichloroethene	0.538	0.605	-12.5	79	0.00
28 T	Bromochloromethane	0.349	0.364	-4.3	70	0.00
29 T	Tetrahydrofuran	0.238	0.253	-6.3	75	0.00
30 C	Chloroform	0.838	0.959	-14.4#	81	0.00
31 T	Cyclohexane	0.737	0.800	-8.5	77	0.00
32 T	1,1,1-Trichloroethane	0.798	0.835	-4.6	74	0.00
33 S	1,2-Dichloroethane-d4	0.510	0.545	-6.9	75	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	73	0.00
35 S	Dibromofluoromethane	0.324	0.319	1.5	73	0.00
36 T	1,1-Dichloropropene	0.421	0.448	-6.4	80	0.00
37 T	Ethyl Acetate	0.472	0.457	3.2	72	0.00
38 T	Carbon Tetrachloride	0.495	0.462	6.7	69	0.00
39 T	Methylcyclohexane	0.547	0.570	-4.2	78	0.00
40 TM	Benzene	1.261	1.364	-8.2	81	0.00
41 T	Methacrylonitrile	0.259	0.255	1.5	76	0.00
42 TM	1,2-Dichloroethane	0.413	0.455	-10.2	82	0.00
43 T	Isopropyl Acetate	0.786	0.744	5.3	71	0.00
44 TM	Trichloroethene	0.382	0.400	-4.7	78	0.00
45 C	1,2-Dichloropropane	0.321	0.337	-5.0#	79	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\WX121418\
 Data File : VX006534.D
 Acq On : 14 Dec 2018 22:45
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 15 05:41:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.231	0.244	-5.6	79	0.00
47 T	Bromodichloromethane	0.448	0.426	4.9	71	0.00
48 T	Methyl methacrylate	0.382	0.383	-0.3	74	0.00
49 T	1,4-Dioxane	0.009	0.010	-11.1	80	0.00
50 S	Toluene-d8	1.178	1.171	0.6	73	0.00
51 T	4-Methyl-2-Pentanone	0.458	0.475	-3.7	76	0.00
52 CM	Toluene	0.830	0.902	-8.7#	82	0.00
53 T	t-1,3-Dichloropropene	0.524	0.474	9.5	67	0.00
54 T	cis-1,3-Dichloropropene	0.544	0.521	4.2	71	0.00
55 T	1,1,2-Trichloroethane	0.337	0.371	-10.1	81	0.00
56 T	Ethyl methacrylate	0.527	0.560	-6.3	79	0.00
57 T	1,3-Dichloropropane	0.533	0.582	-9.2	81	0.00
58 T	2-Chloroethyl Vinyl ether	0.267	0.249	6.7	71	0.00
59 T	2-Hexanone	0.350	0.354	-1.1	75	0.00
60 T	Dibromochloromethane	0.416	0.382	8.2	68	0.00
61 T	1,2-Dibromoethane	0.377	0.405	-7.4	81	0.00
62 S	4-Bromofluorobenzene	0.444	0.460	-3.6	75	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	74	0.00
64 T	Tetrachloroethene	0.377	0.410	-8.8	85	0.00
65 PM	Chlorobenzene	1.061	1.149	-8.3	82	0.00
66 T	1,1,1,2-Tetrachloroethane	0.412	0.401	2.7	74	0.00
67 C	Ethyl Benzene	1.765	1.882	-6.6#	81	0.00
68 T	m/p-Xylenes	0.707	0.755	-6.8	81	0.00
69 T	o-Xylene	0.698	0.738	-5.7	81	0.00
70 T	Styrene	1.130	1.207	-6.8	80	0.00
71 P	Bromoform	0.389	0.311	20.1#	60	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	74	0.00
73 T	Isopropylbenzene	3.453	3.757	-8.8	81	0.00
74 T	N-amyl acetate	1.487	1.399	5.9	71	0.00
75 P	1,1,2,2-Tetrachloroethane	1.105	1.141	-3.3	79	0.00
76 T	1,2,3-Trichloropropane	0.965	1.045	-8.3	75	0.00
77 T	Bromobenzene	0.958	1.017	-6.2	79	0.00
78 T	n-propylbenzene	3.838	4.192	-9.2	81	0.00
79 T	2-Chlorotoluene	2.281	2.482	-8.8	81	0.00
80 T	1,3,5-Trimethylbenzene	2.873	3.109	-8.2	81	0.00
81 T	trans-1,4-Dichloro-2-butene	0.411	0.312	24.1#	57	0.00
82 T	4-Chlorotoluene	2.652	2.857	-7.7	80	0.00
83 T	tert-Butylbenzene	3.065	3.202	-4.5	78	0.00
84 T	1,2,4-Trimethylbenzene	2.967	3.209	-8.2	82	0.00
85 T	sec-Butylbenzene	3.487	3.852	-10.5	82	0.00
86 T	p-Isopropyltoluene	3.140	3.434	-9.4	81	0.00
87 T	1,3-Dichlorobenzene	1.705	1.832	-7.4	80	0.00
88 T	1,4-Dichlorobenzene	1.731	1.815	-4.9	79	0.00
89 T	n-Butylbenzene	2.679	2.863	-6.9	80	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX121418\
Data File : VX006534.D
Acq On : 14 Dec 2018 22:45
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 15 05:41:02 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.640	0.554	13.4	64	0.00
91 T	1,2-Dichlorobenzene	1.692	1.768	-4.5	80	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.283	0.244	13.8	69	0.00
93 T	1,2,4-Trichlorobenzene	1.235	1.263	-2.3	77	0.00
94 T	Hexachlorobutadiene	0.546	0.562	-2.9	76	0.00
95 T	Naphthalene	3.840	4.054	-5.6	80	0.00
96 T	1,2,3-Trichlorobenzene	1.223	1.255	-2.6	77	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6