

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121420\
 Data File : VX019977.D
 Acq On : 14 Dec 2020 11:26
 Operator : JC/MD
 Sample : VX1214WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1214WBL01

Quant Time: Dec 15 00:34:47 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	311807	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.82	114	515557	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.09	117	482115	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	208993	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	195895	47.80	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.60%
35) Dibromofluoromethane	5.46	113	159506	47.87	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.74%
50) Toluene-d8	8.70	98	623707	49.03	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.06%
62) 4-Bromofluorobenzene	11.12	95	220258	45.49	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.98%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121420\
Data File : VX019977.D
Acq On : 14 Dec 2020 11:26
Operator : JC/MD
Sample : VX1214WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1214WBL01

Quant Time: Dec 15 00:34:47 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

