

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121420\
 Data File : VX019979.D
 Acq On : 14 Dec 2020 12:22
 Operator : JC/MD
 Sample : L5052-15
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TU032-32-TB-121020

Quant Time: Dec 15 00:39:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	318079	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	519379	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.09	117	476885	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	208593	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	196852	47.09	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	94.18%		
35) Dibromofluoromethane	5.46	113	160373	47.78	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	95.56%		
50) Toluene-d8	8.70	98	634597	49.52	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	99.04%		
62) 4-Bromofluorobenzene	11.12	95	221693	45.45	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	90.90%		

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121420\
Data File : VX019979.D
Acq On : 14 Dec 2020 12:22
Operator : JC/MD
Sample : L5052-15
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TU032-32-TB-121020

Quant Time: Dec 15 00:39:22 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

