

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121420\
 Data File : VX019980.D
 Acq On : 14 Dec 2020 12:46
 Operator : JC/MD
 Sample : L5059-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CAL-TB01-20201209

Quant Time: Dec 15 00:41:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	321142	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	526874	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.09	117	479363	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	210990	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	201022	47.63	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	95.26%		
35) Dibromofluoromethane	5.46	113	159451	46.83	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	93.66%		
50) Toluene-d8	8.70	98	630027	48.46	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	96.92%		
62) 4-Bromofluorobenzene	11.12	95	223117	45.09	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	90.18%		

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121420\
Data File : VX019980.D
Acq On : 14 Dec 2020 12:46
Operator : JC/MD
Sample : L5059-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CAL-TB01-20201209

Quant Time: Dec 15 00:41:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

