

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121420\
 Data File : VX019983.D
 Acq On : 14 Dec 2020 13:55
 Operator : JC/MD
 Sample : L5052-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TU032-32-5227-121020

Quant Time: Dec 15 14:48:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	305559	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	513309	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.09	117	470285	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	208900	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		6.03	65	195449	48.67	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.34%	
35) Dibromofluoromethane		5.46	113	160286	48.32	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.64%	
50) Toluene-d8		8.70	98	621581	49.08	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.16%	
62) 4-Bromofluorobenzene		11.12	95	224948	46.66	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.32%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	1.18	85	3512	1.255	ug/l	95
3) Chloromethane	1.31	50	2081	0.633	ug/l	93
4) Vinyl Chloride	1.39	62	1856	0.535	ug/l	90
16) Acetone	2.42	43	6510	4.313	ug/l	99
21) trans-1,2-Dichloroethene	3.14	96	3671	1.125	ug/l	99
27) cis-1,2-Dichloroethene	4.56	96	19130	5.008	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121420\
Data File : VX019983.D
Acq On : 14 Dec 2020 13:55
Operator : JC/MD
Sample : L5052-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TU032-32-5227-121020

Quant Time: Dec 15 14:48:18 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

