

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121420\
 Data File : VX019986.D
 Acq On : 14 Dec 2020 15:05
 Operator : JC/MD
 Sample : L5052-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TU032-32-6778-121020

Quant Time: Dec 15 14:50:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	306055	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	511303	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	472666	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	209594	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	198585	49.37	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.74%
35) Dibromofluoromethane	5.46	113	160156	48.47	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.94%
50) Toluene-d8	8.70	98	620106	49.15	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.30%
62) 4-Bromofluorobenzene	11.12	95	221458	46.12	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.24%

Target Compounds

					Qvalue
16) Acetone	2.42	43	1582	1.046	ug/l 99
21) trans-1,2-Dichloroethene	3.14	96	1152	0.352	ug/l 94
27) cis-1,2-Dichloroethene	4.56	96	2554	0.668	ug/l 81
49) 1,4-Dioxane	7.72	88	225	2.490	ug/l # 92
69) o-Xylene	10.68	106	1341	0.221	ug/l 90
73) Isopropylbenzene	11.00	105	3187	0.223	ug/l 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121420\
Data File : VX019986.D
Acq On : 14 Dec 2020 15:05
Operator : JC/MD
Sample : L5052-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TU032-32-6778-121020

Quant Time: Dec 15 14:50:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

