

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121420\
 Data File : VX019992.D
 Acq On : 14 Dec 2020 17:25
 Operator : JC/MD
 Sample : L5022-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WCS-OH-E6

Quant Time: Dec 15 01:55:05 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	331530	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	551795	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.09	117	516298	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	226927	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	208363	47.82	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	95.64%		
35) Dibromofluoromethane	5.46	113	169232	47.45	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	94.90%		
50) Toluene-d8	8.70	98	673542	49.47	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	98.94%		
62) 4-Bromofluorobenzene	11.12	95	241844	46.67	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	93.34%		

Target Compounds

					Qvalue
16) Acetone	2.42	43	21129	12.901	ug/l 98
20) Methylene Chloride	2.83	84	18156	3.534	ug/l 95
43) Isopropyl Acetate	6.41	43	25238	3.083	ug/l 99
95) Naphthalene	13.82	128	63834	4.261	ug/l 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121420\
Data File : VX019992.D
Acq On : 14 Dec 2020 17:25
Operator : JC/MD
Sample : L5022-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WCS-OH-E6

Quant Time: Dec 15 01:55:05 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

