

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121420\
 Data File : VX019994.D
 Acq On : 14 Dec 2020 18:11
 Operator : JC/MD
 Sample : L5065-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 JBD-WC-S-SB-02-0-15

Quant Time: Dec 15 01:59:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	317632	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	526733	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.09	117	493641	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	213444	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	202676	48.55	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	97.10%		
35) Dibromofluoromethane	5.46	113	163820	48.12	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	96.24%		
50) Toluene-d8	8.70	98	657325	50.58	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	101.16%		
62) 4-Bromofluorobenzene	11.12	95	231472	46.79	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	93.58%		

Target Compounds					Qvalue
16) Acetone	2.42	43	15152	9.656 ug/l	92
20) Methylene Chloride	2.83	84	13653	2.459 ug/l	97
43) Isopropyl Acetate	6.41	43	26566	3.400 ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121420\
Data File : VX019994.D
Acq On : 14 Dec 2020 18:11
Operator : JC/MD
Sample : L5065-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
JBD-WC-S-SB-02-0-15

Quant Time: Dec 15 01:59:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

