

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121421\
 Data File : VX025770.D
 Acq On : 14 Dec 2021 11:20
 Operator : JC/MD
 Sample : VX1214MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1214MBL01

Quant Time: Dec 15 07:01:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	148201	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	256336	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	229574	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	98709	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	96497	45.254	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.500%
35) Dibromofluoromethane	5.385	113	83720	46.831	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.660%
50) Toluene-d8	8.647	98	300990	46.210	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.420%
62) 4-Bromofluorobenzene	11.085	95	104738	42.331	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.660%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121421\
Data File : VX025770.D
Acq On : 14 Dec 2021 11:20
Operator : JC/MD
Sample : VX1214MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1214MBL01

Quant Time: Dec 15 07:01:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 14 07:02:11 2021
Response via : Initial Calibration

