

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121421\
 Data File : VX025775.D
 Acq On : 14 Dec 2021 14:35
 Operator : JC/MD
 Sample : M5031-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Dec 15 07:03:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	146984	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	255828	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	232229	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96392	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	97153	45.939	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.880%
35) Dibromofluoromethane	5.391	113	82070	45.999	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.000%
50) Toluene-d8	8.653	98	299087	46.009	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.020%
62) 4-Bromofluorobenzene	11.085	95	113953	46.147	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121421\
Data File : VX025775.D
Acq On : 14 Dec 2021 14:35
Operator : JC/MD
Sample : M5031-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Dec 15 07:03:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 14 07:02:11 2021
Response via : Initial Calibration

