

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121421\
 Data File : VX025780.D
 Acq On : 14 Dec 2021 16:33
 Operator : JC/MD
 Sample : M4988-12 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE122D2-20211207

Quant Time: Dec 15 07:05:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	132565	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	228954	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	207783	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85646	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	89015	46.669	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	93.340%		
35) Dibromofluoromethane	5.385	113	74199	46.469	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	92.940%		
50) Toluene-d8	8.653	98	270865	46.558	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	93.120%		
62) 4-Bromofluorobenzene	11.085	95	103887	47.008	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	94.020%		
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.325	101	1667	1.181	ug/l	94
44) Trichloroethene	7.129	130	631700	351.279	ug/l	95
64) Tetrachloroethene	9.275	164	556	0.335	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121421\
Data File : VX025780.D
Acq On : 14 Dec 2021 16:33
Operator : JC/MD
Sample : M4988-12 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE122D2-20211207

Quant Time: Dec 15 07:05:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 14 07:02:11 2021
Response via : Initial Calibration

