

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121421\
 Data File : VX025781.D
 Acq On : 14 Dec 2021 16:56
 Operator : JC/MD
 Sample : M4988-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE125D2-20211206

Quant Time: Dec 15 07:06:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	134424	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	234835	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	214901	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	89701	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	91755	47.440	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.880%
35) Dibromofluoromethane	5.391	113	76602	46.772	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.540%
50) Toluene-d8	8.653	98	278091	46.603	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.200%
62) 4-Bromofluorobenzene	11.085	95	107527	47.437	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.880%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	21165	14.787	ug/l	98
12) 1,1-Dichloroethene	2.319	96	4979	3.499	ug/l	94
27) cis-1,2-Dichloroethene	4.501	96	6504	3.454	ug/l	92
30) Chloroform	5.099	83	1175	0.372	ug/l	92
44) Trichloroethene	7.129	130	452470	245.311	ug/l	95
64) Tetrachloroethene	9.275	164	9869	5.757	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121421\
Data File : VX025781.D
Acq On : 14 Dec 2021 16:56
Operator : JC/MD
Sample : M4988-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE125D2-20211206

Quant Time: Dec 15 07:06:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 14 07:02:11 2021
Response via : Initial Calibration

