

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121421\
 Data File : VX025784.D
 Acq On : 14 Dec 2021 18:07
 Operator : JC/MD
 Sample : M4988-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE109D3-20211207

Quant Time: Dec 15 07:07:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	133324	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	232653	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	213005	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	88909	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	89890	46.859	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.720%
35) Dibromofluoromethane	5.391	113	75183	46.336	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.680%
50) Toluene-d8	8.653	98	272532	46.100	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.200%
62) 4-Bromofluorobenzene	11.085	95	105685	47.062	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.120%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.325	101	1538	1.083	ug/l	97
44) Trichloroethene	7.129	130	69014	37.768	ug/l	94
64) Tetrachloroethene	9.281	164	1291	0.760	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121421\
Data File : VX025784.D
Acq On : 14 Dec 2021 18:07
Operator : JC/MD
Sample : M4988-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE109D3-20211207

Quant Time: Dec 15 07:07:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 14 07:02:11 2021
Response via : Initial Calibration

