

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121421\
 Data File : VX025785.D
 Acq On : 14 Dec 2021 18:30
 Operator : JC/MD
 Sample : M4988-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE104D3-20211207

Quant Time: Dec 15 07:08:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	136932	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	240235	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	217798	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	91857	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	93646	47.531	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.060%
35) Dibromofluoromethane	5.391	113	77966	46.535	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.080%
50) Toluene-d8	8.653	98	283965	46.518	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.040%
62) 4-Bromofluorobenzene	11.085	95	108059	46.600	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121421\
Data File : VX025785.D
Acq On : 14 Dec 2021 18:30
Operator : JC/MD
Sample : M4988-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE104D3-20211207

Quant Time: Dec 15 07:08:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 14 07:02:11 2021
Response via : Initial Calibration

