

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121421\
 Data File : VX025787.D
 Acq On : 14 Dec 2021 19:17
 Operator : JC/MD
 Sample : M4988-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE104D2-20211207

Quant Time: Dec 15 07:08:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	136479	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	237419	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	214229	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	89726	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	92339	47.023	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	94.040%		
35) Dibromofluoromethane	5.391	113	77688	46.919	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	93.840%		
50) Toluene-d8	8.653	98	282254	46.786	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	93.580%		
62) 4-Bromofluorobenzene	11.085	95	107493	46.906	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	93.820%		
Target Compounds						Qvalue
27) cis-1,2-Dichloroethene	4.489	96	34266	17.922	ug/l	89
30) Chloroform	5.105	83	2882	0.900	ug/l	94
42) 1,2-Dichloroethane	6.104	62	1958	0.769	ug/l	96
44) Trichloroethene	7.129	130	269405	144.471	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121421\
Data File : VX025787.D
Acq On : 14 Dec 2021 19:17
Operator : JC/MD
Sample : M4988-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE104D2-20211207

Quant Time: Dec 15 07:08:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 14 07:02:11 2021
Response via : Initial Calibration

