

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121421\
 Data File : VX025800.D
 Acq On : 15 Dec 2021 01:07
 Operator : JC/MD
 Sample : M4988-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB04-20211207

Quant Time: Dec 15 07:14:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 14 07:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	141389	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	245334	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	220812	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	88620	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	95186	46.790	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.580%
35) Dibromofluoromethane	5.391	113	79471	46.447	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.900%
50) Toluene-d8	8.653	98	289442	46.429	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.860%
62) 4-Bromofluorobenzene	11.085	95	107751	45.502	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.000%
Target Compounds						
					Qvalue	
44) Trichloroethene	7.129	130	1671	0.867	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121421\
Data File : VX025800.D
Acq On : 15 Dec 2021 01:07
Operator : JC/MD
Sample : M4988-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB04-20211207

Quant Time: Dec 15 07:14:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121321W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 14 07:02:11 2021
Response via : Initial Calibration

